



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS


DQ9900AZ.02


LOT 9957B00

Ceveron<sup>®</sup> System Solution, 750mL

REF 9820210

LOT 9957B00

 25-Mar-2015

 31-Mar-2020

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	2.6 – 3.1	2.8	OK
Volumen volume	> 750.0 mL	780.0	OK
Sterilität sterility	0 CFU/mL	0	OK

Durchführung von  
Prepared by

  
Monika WEINBAUER

30-Mar-2015  
Datum / Date:

Freigabe von  
Approved by

  
Lavdrim NASUFI

30-Mar-2015  
Datum / Date:

Ceveron<sup>®</sup> alpha

  
S. SCHWINGENSCHLÖGL

30-Mar-2015  
Datum / Date: