
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9600.01
---	---	-----------

### Ceveron® Cleaning Solution 3% 25 mL

REF      9820300  
LOT      9651B00.01  
      29-Feb-2016

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.2	OK
Volumen Volume	25.0 – 26.5mL	26.0	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, Fibrinogen, Clauss method	TCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1A31B02	2.70 – 4.00g/L	3.46	OK
Usability, Fibrinogen, Clauss method Ceveron® alpha	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	0.93 – 1.39g/L	1.04	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Alessandro BENIGNI

20-Mar-2015  
Datum / Date:

Freigabe von  
Approved by

Lavdrim NASUFI

20-Mar-2015  
Datum / Date: