



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS


DQ9600.01


LOT 9651B00

Ceveron Cleaning Solution 3% 25mL

REF 9820305

LOT 9651B00

 26-Feb-2015

 29-Feb-2016

Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		11.5 – 12.5	12.2	OK	
Volumen Volume		25.0 – 26.5mL	26.0	OK	
Sterilität Sterility		0 CFU/mL	0	OK	
Fibrinogen - aPTT Verschleppung	TCN LOT 1A31B02	Δ sec.	< 5.0%	2.2	OK
		CV	< 5.0%	1.9	OK
Fibrinogen - aPTT carry-over	CCA LOT 3P38B00	Δ sec.	< 5.0%	0.2	OK
		CV	< 5.0%	0.7	OK

Durchführung von
Prepared by


Lucas ROSENMAIER

04-Mär-2015

Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

04-Mär-2015

Datum / Date:

Ceveron® alpha


S. SCHWINGENSCHLÖGL

04-Mär-2015

Datum / Date: