

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9200AZ.02
---	---	-------------

Ceveron® Wash Solution 30mL

REF 9820320
LOT 9231BA0.03
 30-Jun-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	2.0 – 2.6	2.1	OK
Volumen Volume	30.0 – 31.5 mL	30.5	OK
Verwendbarkeit Usability	pH	2.0 – 2.6	2.2 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	


Durchführung von
Prepared by

Lavdrim NASUFI


08-May-2014

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ


08-May-2014

Datum / Date: