


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ920AAZ.02 LOT 9231BA0
-----------------------------------------------------------------------------------	-------------------------------------------------------------	---------------------------------------

Ceveron® Wash Solution, 30mL

REF	9820329
LOT	9231BA0
	12-JUN-2013
	30-JUN-2018

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	2.0 – 2.6	2.1	OK
Volumen volume	30.0 – 31.5mL	30.5	OK
Sterilität sterility	0 CFU/mL	0 CFU	OK

Durchführung von
Prepared by

Lavdrim NASUFI



17. Juni 2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK


17. JUNI 2013
Datum / Date:

Ceveron® alpha

S.SCHWINGENSCHLÖGL


18. Juni 2013
Datum / Date: