

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Antithrombin 1 x 25 IU	PRODUCT #:	81079639
EXP. DATE:	2016-10	LOT #:	N1042832
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654171

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0030	PASS
Intercept @ 37°C	0.40	0.60		0.48	PASS
r ²	0.990	--		0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
Antithrombin, 25 IU	N1042832	2016-10	1 ml

PREPARED BY: N. Patel N.M. Patel DATE: 11/4/2014 11-4-14

APPROVAL SIGNATURE: D. Tencza D. Tencza DATE: 11/4/2014 11/4/14