

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: FXa, 10 x 71 nkat	PRODUCT # : 82098539
EXP. DATE: 2015-01	LOT # : N0720885
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	71	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	0.999	PASS
CONTROL RECOVERY	64.9	71.7	nkat/vial	70.7	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N0720885	2015-01	2 ml

PREPARED BY: NPATEL N.m. Patel

8 24-12
DATE: 8/22/2012

APPROVAL SIGNATURE: DTENCZA J. Dencza

DATE: 8/22/2012 8/24/12