

CERTIFICATE OF ANALYSIS

CHROMOGENIX

| | | | |
|------------|---|------------------|----------|
| PRODUCT: | COATEST APC Resistance Control Plasma Level 1 | PRODUCT #: | 82265063 |
| EXP. DATE: | 2015-01 | LOT #: | N0132663 |
| VENDOR: | INSTRUMENTATION LABORATORY | ASSAYED PER QCM: | C00423 |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | DISPOSITION | |
|-------------------------------------|-----------------------|-----|-------|---------|-------------|------|
| | MIN | MAX | | | | |
| Analysis with COATEST APC R: | | | | | | |
| APT Time | CaCl ₂ | 28 | 34 | seconds | 29 | PASS |
| | APC/CaCl ₂ | 90 | -- | seconds | 110 | PASS |
| APC Ratio | | 3.5 | 4.4 | -- | 3.8 | PASS |

| | | | | | | |
|---------------------------------------|-----------------------|-----|-----|---------|-----|------|
| Analysis with COATEST APC R V: | | | | | | |
| APT Time | CaCl ₂ | -- | 38 | seconds | 36 | PASS |
| | APC/CaCl ₂ | 75 | -- | seconds | 114 | PASS |
| APC Ratio | | 2.7 | 3.2 | -- | 3.2 | PASS |

Kit Components:

| DESCRIPTION | LOT NUMBER | EXP. DATE | VOLUME |
|------------------------|------------|-----------|------------|
| Control Plasma Level 1 | N0132663 | 2015-01 | 1 ml (x 5) |

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:

WPA TEL N.M. Patel

DATE:

1-22-13

APPROVAL SIGNATURE:

DTENCZA J. Dencza

DATE:

1/22/13