

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT #:	82266863
EXP. DATE:	2016-01	LOT #:	N0148040
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
Analysis with COATEST APC R:	MIN	MAX			PASS/FAIL
APT Time <u>CaCl₂</u>	27	44	seconds	40	PASS
<u>APC/CaCl₂</u>	--	101	seconds	80	PASS
APC Ratio	--	2.3	--	2.0	PASS

Analysis with COATEST APC R V:	MIN	MAX	UNITS	RESULTS	DISPOSITION
APT Time <u>CaCl₂</u>	27	40	seconds	35	PASS
APC Ratio	1.4	1.9	--	1.7	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0148040	2016-01	1 ml (x 5)

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N.M. Patel

DATE: 8/25/2014 8:25-14

APPROVAL SIGNATURE: D. Tenca D. Tenca

DATE: 8/25/14
08/25/2014