

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT #:	82266863
EXP. DATE:	2015-07	LOT #:	N0735493
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS		RESULTS	DISPOSITION
Analysis with COATEST APC R:	MIN	MAX	UNITS	PASS/FAIL
APT Time <u>CaCl₂</u>	27	44	seconds	42 PASS
<u>APC/CaCl₂</u>	-	101	seconds	84 PASS
APC Ratio	-	2.3	-	2.0 PASS

Analysis with COATEST APC R V:	MIN	MAX	UNITS	RESULTS	DISPOSITION
APT Time <u>CaCl₂</u>	27	40	seconds	35	PASS
APC Ratio	1.4	1.9	-	1.7	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0735493	2015-07	1 ml (x 5)

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: *N. Patel* *N.M. Patel* DATE: 4/15/2014 *4-15-04*

APPROVAL SIGNATURE: *D. Tenca* *D. Tenca* DATE: 4/15/2014 *4/15/14*