


**DIAPHARIMA PAI-1 Antigen**

# DPGR012A

PAI-1 Antigen

For in vitro research use only

**Example of Calibration curve**



This curve is provided as an example only.

Results must be calculated using only the calibration curve obtained for the series of measurements performed.

**Lot : 140715E – Exp. : 2017-07**

**Standard PAI-1:Ag Lot : 140606M**

PAI-1:Ag concentration: 10.7 ng/ml

**Plasma Control I Lot : 140606N**

PAI-1:Ag concentration: 40 ng/ml

Acceptance range: 34 – 46 ng/ml

**Plasma Control II Lot : 140606O**

PAI-1:Ag concentration: 14 ng/ml

Acceptance range: 11 – 17 ng/ml