

# **ANALYSIS CERTIFICATE**

**HYPHEN BioMed**  
ZAC Neuville Université - 155 rue d'Eragny  
95000 Neuville sur Oise  
France

Tel: +33.1.34.40.65.10  
Fax: +33.1.34.48.72.36  
[www.hyphen-bomed.com](http://www.hyphen-bomed.com)

## **CERTIFICATE OF ANALYSIS**

**ZYMUTEST  
PAI-1 Activity**

**#RK019A**

**Lot : 141211E**

**Expiration date : 2018-01**



## ANALYSIS CERTIFICATE

**PAI-1 Activity Elisa Kit**

**Lot : 141211E**

**CQ Release : 07/01/2015**

**Expiration date : 2018-01**

Components	Volume (mL)	Exp. (months)	Lot #	Exp. date
tPA pre-coated plate	12x8 wells	42	141112D	2018-05
PAI-1 activity Calibrator	3 vials	42	141030E	2018-04
Anti-(h)-PAI-1-HRP immunoconjugate	3 vials	42	141117E	2018-05
F-Sample diluent	2x50	42	141118C	2018-05
Wash solution	1x50	42	140701J	2018-01
Conjugate diluent	1x25	42	140929I	2018-03
PAI-1 Control I	1 vial	42	141121F	2018-05
PAI-1 Control II	1 vial	42	141121G	2018-05
TMB substrate	1x25		140817	2018-08
Sulfuric Acid 0,45M	1x6	42	141104C	2018-05



# ANALYSIS CERTIFICATE

**PAI-1 Activity Elisa Kit**

**Lot : 141211E**

**CQ Release :**

**07/01/2015**

**Expiration date :**

**2018-01**

Analytical data				Specifications	
<b>1. <u>Reactivity</u></b>					
A450 for standard 10 ng/ml :		1,42		> 1.60	
<b>2. <u>Plasma Standard</u></b>					
Concentration		11,7 ng/ml		9-15 ng/ml	
		81,9 IU/ml *		<i>Indicative</i>	
CV		3 %		≤ 10 %	
<small>* 1 ng equivalent to 7 IU (established against WHO International Standard from NIBSC)</small>					
<b>3. <u>Blank value</u></b>					
A450 for sample diluent :		0,003		< 0.100	
SD :		0,002		< 0.015	
<b>4. <u>Calibration curve</u></b>					
Conc.	OD	SD	CV (%)	N	CV (%)
11,7	1,625	0,07	4,5	N = 8	<10
5,85	0,913	0,09	9,3	N = 8	<10
2,93	0,486	0,04	8,3	N = 8	<10
1,17	0,203	0,02	8,5	N = 8	<10
0,59	0,108	0,01	9,7	N = 8	<10
0	0,003	0,002	/	N = 8	<10
<b>5. <u>Detection limit</u></b>					
Sensitivity		Concentration <0,1 ng/ml		≤ 0.5 ng/ml	
<b>6. <u>Controls</u></b>					
	Target value ng/ml	Measured value ng/ml		Acceptancy range ng/ml	
PAI-1 Act. Control I	9,5	9,7		8,1 - 10,9	
PAI-1 Act. Control II	4,0	3,3		3,2 - 4,8	

<p><b>Comments :</b> Accepted under waiver NCI1019</p>	<input checked="" type="checkbox"/> <b>PASSED IN COMPLIANCE</b>
--	---

**Date :** 07/01/2015

**QC Manager :**

A.M.VISSAC



## ZYMUTEST PAI-1 Activity

Référence RK019A

Dosage du PAI-1 actif

Utilisation *in vitro* exclusivement

Uniquement à usage de recherche

## ZYMUTEST PAI-1 Activity

# RK019A

Tissue- Plasminogen Activator Inhibitor Type I Activity

For *in vitro* use only

For research use only

FRANÇAIS

**Lot : 141211E– Exp. : 2018-01**

Calibrator

Lot : 141030E

Taux de PAI-1 actif :

11.7 ng/ml

81.9 IU/ml\*

\* Indicative equivalency established against WHO International Standard from NIBSC

Plasma Control I

Lot :141121F

Taux de PAI-1 actif :

9.5 ng/ml

Intervalle de confiance :

8.1 – 10.9 ng/ml

NE  
13/11/15  
[Signature]

Plasma Control II

Lot : 141121G

Taux de PAI-1 actif :

4.0 ng/ml

Intervalle de confiance :

3.2 – 4.8 ng/ml

ENGLISH

**Lot : 141211E– Exp. : 2018-01**

Calibrator

Lot : 141030E

PAI-1 activity concentration:

11.7 ng/ml

81.9 IU/ml\*

\* Indicative equivalency established against WHO International Standard from NIBSC

Plasma Control I

Lot :141121F

PAI-1 activity concentration:

9.5 ng/ml

Acceptance range:

8.1 -10.9 ng/ml

Plasma Control II

Lot : 141121G

PAI-1 activity concentration:

4.0 ng/ml

Acceptance range:

3.2 – 4.8 ng/ml