



## Certificate of Analysis

**REAADS**  
**VWF Activity ELISA**

Kit Lot: VA-084  
Plate Lot #: VAW-031  
Product #: 10826  
Expiration Date: 2016-03-17

### Quality Control Results

VWF Activity		Testing Date: 2014-10-30		
Attribute	Value	Specification	Result by Assay	
Reference Plasma 100% OD	80 % of Normal	$\geq 1.0$	1.145	
Reagent Background O.D.		$\leq 0.100$	0.018	
Reference Curve R-Squared		$\geq 0.95$	0.98718	
Control Recovery	Label Range	Result by Assay		
Normal Control	78 - 116 %	98.7 %		
Deficient Control	17 - 37 %	29.0 %		

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: **GLADYS REESE**

Signed:   
Title: **QA ASSOCIATE**      2014-11-11

Corgenix, Inc.  
11575 Main Street, Suite 400  
Broomfield, CO 80020  
Phone: (303) 457-4345  
Fax: (303) 457-4519