



Certificate of Analysis

PRODUCT: CoaTest SP Factor FVIII Kit	PRODUCT #: 82408663
EXP. DATE: 2015-03	LOT #: N0333458
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					PASS/FAIL
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.172	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.389	PASS
Standard Curve (r)	0.99	--	--	1.000	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0233032	2016-02	6 ml
CT SP Substrate	N0333453	2015-03	12 ml
Factor Reagent	N0333455	2015-09	10 ml
CT SP Buffer, pH 7.3	N0333377	2015-03	20 ml
CT SP Phospholipid	N0333365	2015-03	2 ml

PREPARED BY: M. Oettinger DATE: 04/12/13

APPROVAL SIGNATURE: D. Teneza DATE: 04/12/13