

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COAMATIC FACTOR VIII	PRODUCT # :	82258563
EXP. DATE:	2015-02	LOT # :	N0333452
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822585/49730503

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.51	--	0.46	PASS
Intercept	--	0.12	--	0.04	PASS
r ²	0.990	--	--	0.999	PASS
LOW RANGE					
Slope	2.8	5.8	--	5.5	PASS
Intercept	--	0.18	--	0.08	PASS
r ²	0.990	--	--	0.999	PASS

Kit Components:

Name	Lot Number	Exp. date
S-2765/ 1-2581	N0333451	2015-03
Factor Reagent	N0233293	2015-02
Buffer	N0333450	2015-03

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:

NPATEL N.m. Patel

DATE:

3-28-13

3/28/2013

APPROVAL SIGNATURE:

DTENCZA D. Tenca

DATE:

3/28/13
3/28/2013