

**CERTIFICATE OF ANALYSIS**

# **CHROMOGENIX**

PRODUCT:	<u>COAMATIC FACTOR VIII</u>	PRODUCT # :	<u>82258563</u>
EXP. DATE:	<u>2015-04</u>	LOT # :	<u>N0534634</u>
VENDOR:	<u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM:	<u>822585/49730503</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.51	--	0.44	PASS
Intercept	--	0.12	--	0.04	PASS
r <sup>2</sup>	0.990	--	--	0.993	PASS
LOW RANGE					
Slope	2.8	5.8	--	4.4	PASS
Intercept	--	0.18	--	0.08	PASS
r <sup>2</sup>	0.990	--	--	0.999	PASS

**Kit Components:**

Name	Lot Number	Exp. date
S-2765/ I-2581	N0434109	2015-04
Factor Reagent	N0534456	2015-05
Buffer	N0534632	2015-05

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: NPATEL DATE: 6/11/2013

APPROVAL SIGNATURE: DTENCZA DATE: 6/11/2013