

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COAMATIC FACTOR VIII	PRODUCT #:	82258563
EXP. DATE:	2016-07	LOT #:	N1043071
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822585/49730503

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.51	--	0.46	PASS
Intercept	--	0.12	--	0.06	PASS
r ²	0.990	--	--	0.996	PASS
LOW RANGE					
Slope	2.8	5.8	--	5.2	PASS
Intercept	--	0.18	--	0.15	PASS
r ²	0.990	--	--	0.999	PASS

Kit Components:

Name	Lot Number	Exp. date
S-2765/1-2581	N0741452	2016-07
Factor Reagent	N1042991	2016-10
Buffer	N0942387	2016-09

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: NPATEL p.m. Patel

DATE: 10-29-14
10/29/2014

APPROVAL SIGNATURE: DTENCZA D. Tenca

DATE: 10/30/14
10/30/2014