

# CHROMOGENIX

## Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT #:	82408663
EXP. DATE:	2015-03	LOT #:	N0333458
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
<b>Coatest SP Factor VIII Kit (82103363)</b>					
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.172	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.389	PASS
Standard Curve (r)	0.99	--	--	1.000	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N0233032	2016-02	6 ml
CT SP Substrate	N0333453	2015-03	12 ml
Factor Reagent	N0333455	2015-09	10 ml
CT SP Buffer, pH 7.3	N0333377	2015-03	20 ml
CT SP Phospholipid	N0333365	2015-03	2 ml

PREPARED BY: M. Oettinger M. Oettinger DATE: 04/12/13 4-12-13

APPROVAL SIGNATURE: D. Teusza P. Pongya DATE: 04/12/13 4/12/13