

CHROMOGENIX

Certificate of Analysis

PRODUCT: CoaTest SP Factor FVIII Kit	PRODUCT #: 82408663
EXP. DATE: 2017-03	LOT #: N0455798
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0092	PASS
Intercept	--	0.250	--	0.102	PASS
Standard Curve (r)	0.99	--	--	0.996	PASS
Low Range					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.110	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0254975	2018-02	6 ml
CT SP Substrate	N0355360	2017-03	12 ml
Factor Reagent	N0355536	2017-09	10 ml
CT SP Buffer, pH 7.3	N0355255	2017-03	20 ml
CT SP Phospholipid	N0355634	2017-03	2 ml

PREPARED BY: N. M. Patel *N. M. Patel* DATE: 04/22/15 *4-22-15*

APPROVAL SIGNATURE: D. Tenca *D. Tenca* DATE: 04/22/15 *4/22/15*