



# Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT # :	82408663
EXP. DATE:	2015-04	LOT # :	N0634944
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					PASS/FAIL
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0099	PASS
Intercept	--	0.250	--	0.170	PASS
Standard Curve (r)	0.99	--	--	0.988	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.243	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N0434006	2016-04	6 ml
CT SP Substrate	N0434107	2015-04	12 ml
Factor Reagent	N0534294	2015-11	10 ml
CT SP Buffer, pH 7.3	N0434008	2015-04	20 ml
CT SP Phospholipid	N0434009	2015-04	2 ml

PREPARED BY: N. Patel *N. m. Patel* DATE: 06/13/13 *6/13-13*

APPROVAL SIGNATURE: D. Tenca *D. Tenca* DATE: 06/13/13 *6/13/13*