

CHROMOGENIX

Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT #:	82408663
EXP. DATE:	2017-06	LOT #:	N0757309
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0104	PASS
Intercept	--	0.250	--	0.125	PASS
Standard Curve (r)	0.99	--	--	0.996	PASS
Low Range					
Slope	0.020	0.060	--	0.047	PASS
Intercept	--	0.406	--	0.117	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0657020	2018-06	6 ml
CT SP Substrate	N0657075	2017-06	12 ml
Factor Reagent	N0657079	2017-12	10 ml
CT SP Buffer, pH 7.3	N0657065	2017-06	20 ml
CT SP Phospholipid	N0656876	2017-06	2 ml

PREPARED BY: TRATEL P. m. b. tel DATE: 07/23/15 7-23-15

APPROVAL SIGNATURE: DTENCZA J. Ruzga DATE: 07/23/15 7/23/15