



Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT #:	82408663
EXP. DATE:	2016-05	LOT #:	N1143180
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0087	PASS
Intercept	--	0.250	--	0.126	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.202	PASS
Standard Curve (r)	0.99	--	--	0.993	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0841692	2017-08	6 ml
CT SP Substrate	N0942242	2016-09	12 ml
Factor Reagent	N1137606	2016-05	10 ml
CT SP Buffer, pH 7.3	N0841953	2016-08	20 ml
CT SP Phospholipid	N0842034	2016-08	2 ml

PREPARED BY: INDIA N. Patel DATE: 11/19/14 11-19-14

APPROVAL SIGNATURE: DTENCZA [Signature] DATE: 11/19/14 11/19/14