

# **CHROMOGENIX**

## Certificate of Analysis

PRODUCT: Coa Test SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2017-06	LOT #: N0757306
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>Coatest SP4 Factor VIII Kit (82409463)</b>					
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0104	PASS
Intercept	--	0.250	--	0.120	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.049	PASS
Intercept	--	0.406	--	0.094	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N0657020	2018-08	(1x) 6 ml
CT SP Substrate	N0657075	2017-06	(1x) 12 ml
Factor Reagent	N0656798	2017-12	(4x) 3 ml
CT SP Buffer, pH 7.3	N0657065	2017-08	(1x) 20 ml
CT SP Phospholipid	N0656876	2017-08	(1x) 2 ml

PREPARED BY: MPATEL N.m. Patel DATE: 07/13/15 7-13-15

APPROVAL SIGNATURE: DTENCZA J. Dencza DATE: 07/13/15 7/13/15