



Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT #:	82409463
EXP. DATE:	2016-08	LOT #:	N0942248
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.119	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.038	PASS
Intercept	--	0.406	--	0.167	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0841692	2017-08	(1x) 6 ml
CT SP Substrate	N0942242	2016-09	(1x) 12 ml
Factor Reagent	N0942247	2017-03	(4x) 3 ml
CT SP Buffer, pH 7.3	N0841953	2016-08	(1x) 20 ml
CT SP Phospholipid	N0842034	2016-08	(1x) 2 ml

PREPARED BY: NPATEL N.M. Patel DATE: 10/21/14 10/21/14

APPROVAL SIGNATURE: DEENBA J. Nonga DATE: 10/21/14 10/21/14

Form QCM0082409463-1, Rev. 02/10