

Certificate of Analysis

Product:

Hirudin Blood Tubes

LOT: A13121C5

Date: 28.01.2014

Parameter

Function testing

The tolerated mean coefficient of variation (CV%) of precision testing of five samples is $\leq 12\%$

Coagulation and vacuum testing

Hirudinized tubes must not show coagulation within 3 hours.

Hirudinized tubes have to be filled to the indicated fill volume.

This batch meets our specification.

Expiry date: 2015-06
Verum Diagnostica GmbH, Munich

Results

ADPtest: 8,9%
ASPItest: 4,1%
COLtest: 6,1%
TRAPtest: 10,5%

corresponds

corresponds



(responsible QC Manager)