





M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-RS 2020	Expiry  2015 - 01
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.102	P-RS 2021	2015-02
M30 Standard B	75 U/L	0.243	P-RS 2022	2015-02
M30 Standard C	150 U/L	0.437	P-RS 2023	2015-02
M30 Standard D	250 U/L	0.722	P-RS 2024	2015-02
M30 Standard E	500 U/L	1.447	P-RS 2025	2015-02
M30 Standard F	750 U/L	2.225	P-RS 2026	2015-02
M30 Standard G	1 000 U/L	2.761	P-RS 2027	2015-02
M5 Coated Microstrips			P-PS 3525	2016-12
M30 Conjugate			P-RS 2031	2015-01
M30 Conjugate Dilution Buffer			P-RS 2030	2015-01
TMB Substrate			P-1811	2016-07
Stop Solution			P-RS 1432	2017-02
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	118	P-RS 2028	2015-02
M30 Control High	550 – 750	660	P-RS 2029	2015-02

Test date: 2014-06-09	Sign: 
Approved by QA: 2014-06-12	Sign: 