




M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-RS 2664	Expiry  2015 - 04
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.091	P-RS 2674	2015-05
M30 Standard B	75 U/L	0.196	P- RS 2675	2015-05
M30 Standard C	150 U/L	0.329	P- RS 2676	2015-05
M30 Standard D	250 U/L	0.590	P- RS 2677	2015-05
M30 Standard E	500 U/L	1.301	P- RS 2678	2015-05
M30 Standard F	750 U/L	1.953	P- RS 2679	2015-05
M30 Standard G	1 000 U/L	2.565	P- RS 2680	2015-05
M5 Coated Microstrips			P-PS 3525	2016-12
M30 Conjugate			P-RS 2684	2015-04
M30 Conjugate Dilution Buffer			P-RS 2683	2015-04
TMB Substrate			P-1811	2016-07
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	130	P-RS 2681	2015-05
M30 Control High	550 – 750	623	P-RS 2682	2015-05

Test date: 2014-09-17	Sign: <i>Gradana Bojelic Bekka</i>
Approved by QA: 2014-09-18	Sign: <i>[Signature]</i>