




M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-RS 3356	Expiry  2015 - 04
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.080	P-RS 2674	2015-05
M30 Standard B	75 U/L	0.177	P-RS 2675	2015-05
M30 Standard C	150 U/L	0.299	P-RS 2676	2015-05
M30 Standard D	250 U/L	0.572	P-RS 2677	2015-05
M30 Standard E	500 U/L	1.259	P-RS 2678	2015-05
M30 Standard F	750 U/L	1.954	P-RS 2679	2015-05
M30 Standard G	1 000 U/L	2.439	P-RS 2680	2015-05
M5 Coated Microstrips			P-RS 3255	2017-10
M30 Conjugate			P-RS 2684	2015-04
M30 Conjugate Dilution Buffer			P-RS 2683	2015-04
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	135	P-RS 2681	2015-05
M30 Control High	550 – 750	640	P-RS 2682	2015-05

Test date: 2014-10-15	Sign: <i>Gerdyana Bjelic Bekar</i>
Approved by QA: 2014-10-16	Sign: <i>[Signature]</i>