





M30 Apoptosense[®] ELISA

REF 10011

Kit Lot No. LOT PE-RS 3464	Expiry  2015-06
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.084	P-RS 3454	2015-08
M30 Standard B	75 U/L	0.213	P-RS 3455	2015-08
M30 Standard C	150 U/L	0.382	P-RS 3456	2015-08
M30 Standard D	250 U/L	0.703	P-RS 3457	2015-08
M30 Standard E	500 U/L	1.512	P-RS 3458	2015-08
M30 Standard F	750 U/L	2.279	P-RS 3459	2015-08
M30 Standard G	1 000 U/L	2.722	P-RS 3460	2015-08
M5 Coated Microstrips			P-RS 3255	2017-10
M30 Conjugate			P-RS 3463	2015-06
M30 Conjugate Dilution Buffer			P-RS 3453	2015-07
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-RS 2578	2017-02
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	124	P-RS 3461	2015-08
M30 Control High	550 – 750	629	P-RS 3462	2015-08

Test date: 2014-11-27	Sign: 
Approved by QA: 2014-11-28	Sign: 