





M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-RS 3594	Expiry  2015-08
-----------------------------------------	----------------------------------------------------------------------------------------------------------

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.069	P-RS 3604	2015-10
M30 Standard B	75 U/L	0.159	P-RS 3605	2015-10
M30 Standard C	150 U/L	0.318	P-RS 3606	2015-10
M30 Standard D	250 U/L	0.568	P-RS 3607	2015-10
M30 Standard E	500 U/L	1.245	P-RS 3608	2015-10
M30 Standard F	750 U/L	1.940	P-RS 3609	2015-10
M30 Standard G	1 000 U/L	2.508	P-RS 3610	2015-10
M5 Coated Microstrips			P-RS 3586	2017-11
M30 Conjugate			P-RS 3613	2015-08
M30 Conjugate Dilution Buffer			P-RS 3614	2015-08
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-RS 2578	2017-02
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	126	P-RS 3611	2015-10
M30 Control High	550 – 750	620	P-RS 3612	2015-10

Test date: 2015-01-22	Sign: 
Approved by QA: 2015-01-23	Sign: 