




M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT PE-SS 1823	Expiry  2016-01
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.098	P-SS 1835	2016-02
M30 Standard B	75 U/L	0.218	P-SS 1836	2016-02
M30 Standard C	150 U/L	0.401	P-SS 1837	2016-02
M30 Standard D	250 U/L	0.692	P-SS 1838	2016-02
M30 Standard E	500 U/L	1.481	P-SS 1839	2016-02
M30 Standard F	750 U/L	2.107	P-SS 1840	2016-02
M30 Standard G	1 000 U/L	2.696	P-SS 1841	2016-02
M5 Coated Microstrips			P-RS 3586	2017-11
M30 Conjugate			P-SS 1834	2016-01
M30 Conjugate Dilution Buffer			P-SS 1824	2016-01
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-SS 1301	2017-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	125	P-SS 1842	2016-02
M30 Control High	550 – 750	678	P-SS 1843	2016-02

Test date: 2015-06-16	Sign: <i>Gordana Bjelic Beba</i>
Approved by QA: 2015-06-18	Sign: <i>Slava Muzijic</i>