




M30 Apoptosense[®] ELISA

REF 10011

Kit Lot No. LOT PE-SS 1923	Expiry  2016-04
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.122	P-SS 2479	2016-05
M30 Standard B	75 U/L	0.245	P-SS 2480	2016-05
M30 Standard C	150 U/L	0.433	P-SS 2481	2016-05
M30 Standard D	250 U/L	0.737	P-SS 2482	2016-05
M30 Standard E	500 U/L	1.634	P-SS 2483	2016-05
M30 Standard F	750 U/L	2.201	P-SS 2484	2016-05
M30 Standard G	1 000 U/L	2.939	P-SS 2485	2016-05
M5 Coated Microstrips			P-SS 1878	2018-06
M30 Conjugate			P-SS 2469	2016-04
M30 Conjugate Dilution Buffer			P-SS 2371	2016-04
TMB Substrate			P-SS 2272	2018-11
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-RS 2578	2017-02
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	115	P-SS 2486	2016-05
M30 Control High	550 – 750	628	P-SS 2487	2016-05

Test date: 2015-09-14	Sign: <i>Gerdiana Bjelic Bekar</i>
Approved by QA: 2015-03-17	Sign: <i>Anne Muzylka</i>