




M30 Apoptosense[®] ELISA

REF 10011

Kit Lot No. LOT PE-SS 2998	Expiry  2016-07
---	--

Component		A _{450 nm}	LOT	
M30 Standard A	0 U/L	0.097	P-SS 3015	2016-10
M30 Standard B	75 U/L	0.200	P-SS 3016	2016-10
M30 Standard C	150 U/L	0.360	P-SS 3017	2016-10
M30 Standard D	250 U/L	0.625	P-SS 3018	2016-10
M30 Standard E	500 U/L	1.391	P-SS 3019	2016-10
M30 Standard F	750 U/L	2.211	P-SS 3020	2016-10
M30 Standard G	1 000 U/L	2.835	P-SS 3021	2016-10
M5 Coated Microstrips			P-SS 1878	2018-06
M30 Conjugate			P-SS 2999	2016-09
M30 Conjugate Dilution Buffer			P-SS 3001	2016-07
TMB Substrate			P-SS 2272	2018-11
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-SS 1301	2017-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M30 Control Low	100 – 150	123	P-SS 3022	2016-10
M30 Control High	550 – 750	638	P-SS 3023	2016-10

Test date: 2016-01-13	Sign: <i>Gardara Björk Pólsen</i>
Approved by QA: 2016-01-14	Sign: <i>Alice Kuzko</i>

ASSAYS

M30 Apoptosense® ELISA

Prod. No. 10011

M65® ELISA

Prod. No. 10020

M65 EpiDeath® ELISA

Prod. No. 10040

M30 CytoDeath™ ELISA

Prod. No. 10900

ANTIBODIES

M30 CytoDEATH™

- | | |
|----------------|-----------------|
| ▪ Unconjugated | Prod. No. 10700 |
| ▪ Biotin | Prod. No. 10750 |
| ▪ Fluorescein | Prod. No. 10800 |
| ▪ Orange | Prod. No. 10830 |

M5 Keratin 18

Prod. No. 10600

M6 Keratin 18

Prod. No. 10650

Certificate of Analysis

M30 Apoptosense® ELISA

REF 10011