





M65[®] ELISA

REF 10020

Kit Lot No. LOT PE-RS 1177	Expiry  2015-01
---	--

Component		A _{450 nm}	LOT	
M65 Standard A	0 U/L	0.076	P-RS 1089	2016-12
M65 Standard B	125 U/L	0.217	P-RS 1090	2015-01
M65 Standard C	250 U/L	0.373	P-RS 1091	2015-01
M65 Standard D	500 U/L	0.654	P-RS 1092	2015-01
M65 Standard E	750 U/L	0.997	P-RS 1093	2015-01
M65 Standard F	1 200 U/L	1.506	P-RS 1094	2015-01
M65 Standard G	2 000 U/L	2.448	P-RS 1095	2015-01
M65 Coated Microstrips			P-PS 3557	2016-12
M65 HRP Conjugate			P-RS 1260	2015-07
M65 Conjugate Dilution Buffer			P-RS 1104	2015-09
TMB Substrate			P-1811	2016-07
Stop Solution			P-PS 3596	2016-12
Wash Solution			P-1925	2016-01
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M65 Control Low	160 – 240	197	P-RS 1096	2015-01
M65 Control High	850 – 1 150	1018	P-RS 1097	2015-01

Test date: 2014-03-28	Sign: 
Approved by QA: 2014-04-01	Sign: 