




M65[®] ELISA

REF 10020

Kit Lot No. LOT PE-RS 2393	Expiry  2015-07
---	--

Component		A _{450 nm}	LOT	
M65 Standard A	0 U/L	0.066	P-RS 2605	2015-07
M65 Standard B	125 U/L	0.179	P-RS 2929	2015-08
M65 Standard C	250 U/L	0.286	P-RS 2930	2015-08
M65 Standard D	500 U/L	0.569	P- RS 2931	2015-08
M65 Standard E	750 U/L	0.851	P- RS 2932	2015-08
M65 Standard F	1 200 U/L	1.358	P- RS 2933	2015-08
M65 Standard G	2 000 U/L	2.231	P- RS 2934	2015-08
M6 Coated Microstrips			P-PS 3557	2016-12
M65 HRP Conjugate			P-RS 2633	2015-10
M65 Conjugate Dilution Buffer			P-RS 2430	2016-02
TMB Substrate			P-1811	2016-07
Stop Solution			P-RS 1432	2017-02
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M65 Control Low	160 – 240	199	P- RS 2935	2015-08
M65 Control High	850 – 1 150	990	P- RS 2936	2015-08

Test date: 2014-09-19	Sign: <i>Gordana Prjelic Bobba</i>
Approved by QA: <i>2014-09-25</i>	Sign: <i>[Signature]</i>