




# M65<sup>®</sup> ELISA

**REF** 10020

Kit Lot No. <b>LOT</b> PE-RS 3566	Expiry  2016-03
---	--

Component		A <sub>450 nm</sub>	<b>LOT</b>	
M65 Standard A	0 U/L	0.055	P-RS 3727	2016-03
M65 Standard B	125 U/L	0.140	P-RS 3728	2016-03
M65 Standard C	250 U/L	0.246	P-RS 3729	2016-03
M65 Standard D	500 U/L	0.529	P-RS 3730	2016-03
M65 Standard E	750 U/L	0.834	P-RS 3731	2016-03
M65 Standard F	1 200 U/L	1.392	P-RS 3732	2016-03
M65 Standard G	2 000 U/L	2.246	P-RS 3733	2016-03
M6 Coated Microstrips			P-RS 3543	2017-11
M65 HRP Conjugate			P-RS 3717	2016-04
M65 Conjugate Dilution Buffer			P-RS 3716	2016-08
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-RS 2578	2017-02
Controls	Expected value (U/L)	Obtained value (U/L)	<b>LOT</b>	
M65 Control Low	160 – 240	202	P-RS 3734	2016-03
M65 Control High	850 – 1 150	943	P-RS 3735	2016-03

Test date: 2015-03-17	Sign: <i>Gordana Bijelic Belca</i>
Approved by QA: 2015-03-18	Sign: <i>[Signature]</i>