




M65[®] ELISA

REF 10020

Kit Lot No. LOT PE-SS 1855	Expiry  2016-07
---	--

Component		A _{450 nm}	LOT	
M65 Standard A	0 U/L	0.060	P-SS 1867	2016-07
M65 Standard B	125 U/L	0.130	P-SS 1868	2016-07
M65 Standard C	250 U/L	0.252	P-SS 1869	2016-07
M65 Standard D	500 U/L	0.503	P-SS 1870	2016-07
M65 Standard E	750 U/L	0.804	P-SS 1871	2016-07
M65 Standard F	1 200 U/L	1.268	P-SS 1872	2016-07
M65 Standard G	2 000 U/L	2.164	P-SS 1873	2016-07
M6 Coated Microstrips			P-SS 1879	2018-05
M65 HRP Conjugate			P-SS 3102	2017-02
M65 Conjugate Dilution Buffer			P-SS 1856	2017-01
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-SS 1301	2017-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M65 Control Low	160 – 240	172	P-SS 1874	2016-07
M65 Control High	850 – 1 150	908	P-SS 1875	2016-07

Test date: 2015-10-27	Sign: <i>Gordana Bjelic Belca</i>
Approved by QA: 2015-10-29	Sign: <i>Durkin</i>