




# M65<sup>®</sup> ELISA

**REF** 10020

Kit Lot No. <b>LOT</b> PE-SS 3378	Expiry  2017-01
---	--

Component		A <sub>450 nm</sub>	LOT	
M65 Standard A	0 U/L	0.064	P-SS 3390	2017-01
M65 Standard B	125 U/L	0.156	P-SS 3391	2017-01
M65 Standard C	250 U/L	0.281	P-SS 3392	2017-01
M65 Standard D	500 U/L	0.534	P-SS 3393	2017-01
M65 Standard E	750 U/L	0.786	P-SS 3394	2017-01
M65 Standard F	1 200 U/L	1.249	P-SS 3395	2017-01
M65 Standard G	2 000 U/L	2.131	P-SS 3396	2017-01
M6 Coated Microstrips			P-SS 1879	2018-05
M65 HRP Conjugate			P-SS 3380	2017-02
M65 Conjugate Dilution Buffer			P-SS 3379	2017-08
TMB Substrate			P-SS 2272	2018-11
Stop Solution			P-TS 1034	2019-01
Wash Tablet			P-SS 1301	2017-08
Controls	Expected value (U/L)	Obtained value (U/L)	LOT	
M65 Control Low	160 – 240	216	P-SS 3397	2017-01
M65 Control High	850 – 1 150	1025	P-SS 3398	2017-01

Test date: 2016-02-08	Sign: 
Approved by QA: 2016-02-10	Sign: 

ASSAYS

---

**M30 Apoptosense® ELISA**

Prod. No. 10011

**M65® ELISA**

Prod. No. 10020

**M65 EpiDeath® ELISA**

Prod. No. 10040

**M30 CytoDeath™ ELISA**

Prod. No. 10900

ANTIBODIES

---

**M30 CytoDEATH™**

- Unconjugated                      Prod. No. 10700
- Biotin                                Prod. No. 10750
- Fluorescein                        Prod. No. 10800
- Orange                                Prod. No. 10830

**M5 Keratin 18**

Prod. No. 10600

**M6 Keratin 18**

Prod. No. 10650

**Certificate of Analysis****M65® ELISA****REF 10020**