





# M65 EpiDeath® ELISA

**REF** 10040

Kit Lot No. <b>LOT</b> PE-RS 1178	Expiry  2015-09
---	--

Component		A <sub>450 nm</sub>	LOT	
M65 EpiDeath Standard A	0 U/L	0.051	P-RS 1384	2016-12
M65 EpiDeath Standard B	200 U/L	0.155	P-RS 1385	2015-09
M65 EpiDeath Standard C	400 U/L	0.278	P-RS 1386	2015-09
M65 EpiDeath Standard D	800 U/L	0.556	P-RS 1387	2015-09
M65 EpiDeath Standard E	1 200 U/L	0.782	P-RS 1388	2015-09
M65 EpiDeath Standard F	2 000 U/L	1.299	P-RS 1389	2015-09
M65 EpiDeath Standard G	3 000 U/L	1.899	P-RS 1390	2015-09
M65 EpiDeath Standard H	5 000 U/L	2.678	P-RS 1391	2015-09
M5 Coated Microstrips			P-PS 3525	2016-12
M65 EpiDeath Conjugate			P-RS 1612	2016-03
M65 EpiDeath Conjugate Dilution Buffer			P-RS 1395	2016-03
TMB Substrate			P-1811	2016-07
Stop Solution			P-RS 1432	2017-02
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value U/L	Obtained value U/L	LOT	
M65 EpiDeath Control Low	240 – 360	309	P-RS 1392	2015-09
M65 EpiDeath Control High	2 000 – 3 000	2411	P-RS 1393	2015-09

Test date: 2014-04-02	Sign: 
Approved by QA: 2014-04-04	Sign: 