





M65 EpiDeath® ELISA

REF 10040

Kit Lot No. LOT PE-RS 2657	Expiry  2015-12
---	--

Component		A _{450 nm}	LOT	
M65 EpiDeath Standard A	0 U/L	0.045	P-RS 2639	2016-05
M65 EpiDeath Standard B	200 U/L	0.153	P-RS 2640	2016-05
M65 EpiDeath Standard C	400 U/L	0.270	P-RS 2641	2016-05
M65 EpiDeath Standard D	800 U/L	0.525	P-RS 2642	2016-05
M65 EpiDeath Standard E	1 200 U/L	0.723	P-RS 2643	2016-05
M65 EpiDeath Standard F	2 000 U/L	1.199	P-RS 2644	2016-05
M65 EpiDeath Standard G	3 000 U/L	1.686	P-RS 2645	2016-05
M65 EpiDeath Standard H	5 000 U/L	2.864	P-RS 2646	2016-05
M5 Coated Microstrips			P-RS 3255	2017-10
M65 EpiDeath Conjugate			P-RS 2638	2016-07
M65 EpiDeath Conjugate Dilution Buffer			P-RS 2698	2016-08
TMB Substrate			P-RS 3054	2018-04
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-PS 3613	2015-12
Controls	Expected value U/L	Obtained value U/L	LOT	
M65 EpiDeath Control Low	240 – 360	306	P-RS 2647	2016-05
M65 EpiDeath Control High	2 000 – 3 000	2365	P-RS 2648	2016-05

Test date: 2014-11-10	Sign: 
Approved by QA: 2014-11-10	Sign: 