

ASSAYS

---

**M30 Apoptosense® ELISA**

Prod. No. 10011

**M65® ELISA**

Prod. No. 10020

**M65 EpiDeath® ELISA**

Prod. No. 10040

**M30 CytoDeath™ ELISA**

Prod. No. 10900

ANTIBODIES

---

**M30 CytoDEATH™**

- Unconjugated                      Prod. No. 10700
- Biotin                                      Prod. No. 10750
- Fluorescein                              Prod. No. 10800
- Orange                                      Prod. No. 10830

**M5 Keratin 18**

Prod. No. 10600


**M6 Keratin 18**



Prod. No. 10650

**Certificate of Analysis****M65 EpiDeath® ELISA****REF 10040**

# M65 EpiDeath® ELISA

**REF** 10040

Kit Lot No.  <b>LOT</b>  PE-SS 2923	Expiry    2017-02
---	--

Component		A <sub>450 nm</sub>	LOT	
M65 EpiDeath Standard A	0 U/L	0.051	P-SS 2936	2017-07
M65 EpiDeath Standard B	200 U/L	0.162	P-SS 2937	2017-07
M65 EpiDeath Standard C	400 U/L	0.295	P-SS 2938	2017-07
M65 EpiDeath Standard D	800 U/L	0.612	P-SS 2939	2017-07
M65 EpiDeath Standard E	1 200 U/L	0.888	P-SS 2940	2017-07
M65 EpiDeath Standard F	2 000 U/L	1.367	P-SS 2941	2017-07
M65 EpiDeath Standard G	3 000 U/L	1.977	P-SS 2942	2017-07
M65 EpiDeath Standard H	5 000 U/L	2.868	P-SS 2943	2017-07
M5 Coated Microstrips			P-SS 1878	2018-06
M65 EpiDeath Conjugate			P-SS 2925	2017-10
M65 EpiDeath Conjugate Dilution Buffer			P-SS 3273	2017-12
TMB Substrate			P-SS 2272	2018-11
Stop Solution			P-RS 2738	2017-07
Wash Tablet			P-RS 2578	2017-02
Controls	Expected value U/L	Obtained value U/L	LOT	
M65 EpiDeath Control Low	240 – 360	335	P-SS 2944	2017-07
M65 EpiDeath Control High	2 000 – 3 000	2 459	P-SS 2945	2017-07

Test date: 2016-01-25	Sign: <i>Godina Bjelic Bekar</i>
Approved by QA: 2016-01-26	Sign: <i>Ane Maryko</i>