

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2302, 25 mg Substrate	PRODUCT #: 00082034039
EXP. DATE: 2015-12	LOT #: N0821229
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.6	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0821229	2015-12	3 mL

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 11-Sep-2012 *9-11-12*

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 11-Sep-2012 *9/12/12*

Form QCM00820340-2 Rev. 03/09