

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

PRODUCT:	S-2288, 25 mg Substrate	PRODUCT #:	82085239
EXP. DATE:	2018-10	LOT #:	N1059102
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820852

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	26.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2288	N1059102	2018-10	2 ml

PREPARED BY: M. Oettinger *M. Oettinger*

DATE: 11/10/2015 *11-10-15*

APPROVAL SIGNATURE: D. Tencza *D. Tencza*

DATE: 11/10/2015