

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: <u>S-2765, 25 mg Substrate</u>	PRODUCT # : <u>82141339</u>
EXP. DATE: <u>2018-07</u>	LOT # : <u>N0757345</u>
VENDOR: <u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM: <u>821413</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2765	N0757345	2018-07	3 ml

PREPARED BY:

M. Oettinger

DATE:

7/22/2015

APPROVAL SIGNATURE:

DATE: