

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

PRODUCT: S-2765, 25 mg Substrate	PRODUCT # : 82141339
EXP. DATE: 2015-08	LOT # : N0821104
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 821413

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.1	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2765	N0821104	2015-08	3 ml

PREPARED BY: N. Patel *N. Patel* DATE: 8/27/2012 *8-28-12*

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 8/27/2012 *8/28/12*