

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2403, 25 mg Substrate	PRODUCT # : 82225439
EXP. DATE: 2015-01	LOT # : N0128972
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 822254

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.6	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2403	N0128972	2015-01	3 ml

PREPARED BY: NPATEL N.m. Patel

DATE: 2-13-12
2/13/2012

APPROVAL SIGNATURE: DTENCZA J. Dencza

DATE: 2/13/12
2/13/2012