



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRL00AZ

t-PA Antigen EDTA ELISA Kit 96T

REF TC 12007
LOT RL34B00.01
 31 - Mar -2015

Inhalt Content	Reagenz Reagent	LOT	
1x	GN-Anti t-PA Plate	RGN31B0	31-Mar-2015
1x0.3mL	KM-POX anti t-PA Conjugate	RKM32B0	30-Jun-2015
1x80mL	Wash buffer Concentrate	RNA33B0	31-May-2015
1x90mL	Incubation buffer	RNB32B0	30-Apr-2015
1x12mL	TMB Substrat 12mL	RKN32B0	31-May-2015
1x12mL	Stop solution	RKK32B0	30-Jun-2015
1x20mL	Dilution buffer	RAD31B0	31-May-2015
1x0.5mL	Calibrator 1 t-PA	1AV33B0	31-May-2015
1x0.5mL	Calibrator 2 t-PA	2AV32B0	31-May-2015
1x0.5mL	Calibrator 3 t-PA	3AV32B0	31-Jul-2015
1x0.5mL	Calibrator 4 t-PA	4AV32B0	31-Jul-2015
1x0.5mL	Calibrator 5 t-PA	5AV31B0	31-May-2015
1x0.5mL	Control High t-PA	AAV32B0	31-Jul-2015
1x0.5mL	Control Low t-PA	BAV32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	1.0 – 2.5	1.612	OK
	Calibrator 5	< 0.1	0.005	OK
Verwendbarkeit Usability	High Control (AA32B0)	13.52 – 22.53 ng/mL	17.185	OK
	Low Control (BAV32B0)	2.68 – 4.46 ng/mL	3.861	OK

Durchführung von
Prepared by

Michael RIHA

15. Juli 2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK

16. JULI 2013
Datum / Date: