

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRL00AZ
---	---	----------

t-PA Antigen EDTA ELISA Kit

REF TC 12007
LOT RL35B00.01
 30 - Apr -2015

Inhalt Content	Reagenz Reagent	LOT	
1x	GN-Anti t-PA Plate	RGN32B0	31-Jul-2015
1x0.3mL	KM-POX anti t-PA Conjugate	RKM32B0	30-Jun-2015
1x80mL	Wash buffer Concentrate	RNA33B0	31-May-2015
1x90mL	Incubation buffer	RNB32B0	30-Apr-2015
1x12mL	TMB Substrat 12mL	RKN32B0	31-May-2015
1x12mL	Stop solution	RKK32B0	30-Jun-2015
1x20mL	Dilution buffer	RAD31B0	31-May-2015
1x0.5mL	Calibrator 1 t-PA	1AV33B0	31-May-2015
1x0.5mL	Calibrator 2 t-PA	2AV32B0	31-May-2015
1x0.5mL	Calibrator 3 t-PA	3AV32B0	31-Jul-2015
1x0.5mL	Calibrator 4 t-PA	4AV32B0	31-Jul-2015
1x0.5mL	Calibrator 5 t-PA	5AV31B0	31-May-2015
1x0.5mL	Control High t-PA	AAV32B0	31-Jul-2015
1x0.5mL	Control Low t-PA	BAV32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.710	OK
	Calibrator 5	< 0.1	0.014	OK
Verwendbarkeit Usability	High Control (AAV91A1)	(13.52 – 22.53)ng/mL	18.551	OK
	Low Control (BAV01B0)	(2.68 – 4.46)ng/mL	4.020	OK

Durchführung von
Prepared by

Michael RIHA



20-Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER



23. SEP. 2013
Datum / Date: