



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRL00AZ

## t-PA Antigen EDTA ELISA Kit

REF      TC12007  
LOT      RL42B00.01  
      31-May-2015

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x	GN-Anti t-PA Plate	RGN41B0	30-Jun-2016
1x0.3mL	KM-POX anti t-PA Conjugate	RKM32B0	30-Jun-2015
1x80mL	Wash buffer Concentrate	RNA41B0	29-Feb-2016
1x90mL	Incubation buffer	RNB42B0	30-Apr-2016
1x12mL	TMB Substrat 12mL	RKN34B0	30-Nov-2015
1x12mL	Stop solution	RKK41B0	31-Mar-2016
1x20mL	Dilution buffer	RAD31B0	31-May-2015
1x0.5mL	Calibrator 1 t-PA	1AV33B0	31-May-2015
1x0.5mL	Calibrator 2 t-PA	2AV32B0	31-May-2015
1x0.5mL	Calibrator 3 t-PA	3AV32B0	31-Jul-2015
1x0.5mL	Calibrator 4 t-PA	4AV32B0	31-Jul-2015
1x0.5mL	Calibrator 5 t-PA	5AV31B0	31-May-2015
1x0.5mL	Control High t-PA	AAV32B0	31-Jul-2015
1x0.5mL	Control Low t-PA	BAV32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	2.23	OK
	Calibrator 5	< 0.1	0.03	OK
Verwendbarkeit Usability	High Control (AAV32B0)	13.52 – 22.53 ng/mL	17.16	OK
	Low Control (BAV32B0)	2.68 – 4.46 ng/mL	3.98	OK

Durchführung von  
Prepared by

25-Sep-2014  
Datum / Date:

Freigabe von  
Approved by

**Michael RIHA**

8-SEP-2014  
Datum / Date: