

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRXXAZ.01
---	---	------------

Fibronectin ELISA Kit 96T

REF TC 12030
LOT 120301B4.01
 30 - Apr - 2015

Inhalt Content	Reagenz Reagent	LOT	
1x	anti Fibronectin Plate	RGW32B0	31-Dec-2015
1x0.3mL	POX anti Fibronectin	RAG32B0	31-Dec-2015
1 x 20 ml	Wash buffer	RBE31B0	30-Apr-2015
3 x 20 ml	Dilution buffer	RDS31B0	31-Jul-2015
1x12mL	TMB Substrat	RKN33B0	30-Sep-2015
1x12mL	Stop solution	RKK33B0	31-Oct-2015
1 x 0,6 ml	Standard Fibronectin	RAJ81A4	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1	0.93*	OK
	Calibrator 6	< 0.1	0.05	OK
Wiederfindung der Kontrollen/ Recovery of controls	Internal control 1	70 – 148 µg/mL	61*	OK
	Internal control 2	70 – 148 µg/mL	62.1*	OK

* Sonderfreigabe / special release

Durchführung von
Prepared by

Fikret NASUFI 

03-Apr-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA 

03 APR - 2014
Datum / Date: